TL LABORATORIES

Certificate of Analysis

T6I1151



JAG Alliance, LLC
Client Lic#:
83 Knight Boxx Rd Orange Park, FL
32065
JAG Alliance, LLC

Compliance for Retail: **PASS**

1000mg CBN Tincture Drop (MCT) - Mint - 1oz/ 30ml

Sample ID: T6I1151-01 CCB ID:Lot E258611546-1
Matrix: Derivative Non-inhalable

Total Sample Received: 30ml
Total Units Received: 1
Unit Weight: 30g

Date Sampled: 09/24/2026
Date Received: 09/24/2026
Date Reported: 09/29/2026

Safety Summary

Foreign Materials	Heavy Metals	Microbials	Moisture Content	Mycotoxins
PASS	PASS	PASS	NOT TESTED	PASS
Pesticides	**Residual Solvents**	**Terpenes**	**Water Activity**	**Vitamins Supplements**
PASS	PASS	TESTED	PASS	NOT TESTED

Potency Summary

Cannabinoids

Date Prepared: 09/25/2026 08:36 Prep ID: 4599 Specimen Wt: 0.108 g
Date Analyzed: 09/25/2026 22:00 Analyst ID: SYSTEM (SY Instrument: HPLC VWD
Lab Batch: 2639124 Prep/Analysis Method: SOP 1357

Analyte	Dilution	Results	mg/Serving	mg/Unit
		%	mg	mg
Cannabichromene (CBC)	10	ND	N/A	ND
Cannabidiol (CBD)	10	ND	N/A	ND
Cannabidiolic acid (CBDA)	10	ND	N/A	ND
Cannabidivarin (CBDV)	10	ND	N/A	ND
Cannabigerol (CBG)	10	ND	N/A	ND
Cannabigerolic acid (CBGA)	10	ND	N/A	ND
Cannabinol (CBN)	10	3.79	N/A	1,136
d8 - Tetrahydrocannabinoid (d8-THC)	10	ND	N/A	ND
d9 - Tetrahydrocannabinoid (d9-THC)	10	ND	N/A	ND
d9 - Tetrahydrocannabinolic acid (THCA)	10	ND	N/A	ND
Tetrahydrocannabivarin (THCV)	10	ND	N/A	ND
Total THC		0.0000	0.000	0.000
Total CBD		0.0000	0.000	0.000
Total Cannabinoids		3.79	0.000	1,136

Serving Size: g

Total THC= THCa * 0.877 + d9-THC. Total CBD= CBDa * 0.877 + CBD.
mg/Unit = Unit Weight g * Total THC/CBD mg/g
LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

TL Laboratories
10350 Fisher Ave, Tampa, Florida 33619
813-731-2451 / www.terplifelabs.com

DEA Registration Number: RT0720878

Brian C. Spann
Laboratory Director

PJLA Testing
Accreditation #116557

The data contained therein are based on sound scientific analytical procedures and judgment. TL Laboratories strives to deliver high quality results. This report shall not be reproduced without written consent from TL Laboratories The results of this report relate only to the material or product received and analyzed. Test Results are confidential unless explicitly waived otherwise. This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISOILAC-IAF Communiqué dated April 2017)

TL LABORATORIES

Certificate of Analysis

T6I1151



JAG Alliance, LLC
Client Lic#:
83 Knight Boxx Rd Orange Park, FL
32065
JAG Alliance, LLC

Compliance for Retail: **PASS**

1000mg CBN Tincture Drop (MCT) - Mint - 1oz/ 30ml

Sample ID: T6I1151-01 CCB ID:Lot E258611546-1
Matrix: Derivative Non-inhalable

Total Sample Received: 30ml
Total Units Received: 1
Unit Weight: 30g

Date Sampled: 09/24/2026
Date Received: 09/24/2026
Date Reported: 09/29/2026

Pesticides

Pass

Prep Mehtod: SOP 1363	Batch	Instrument	Analyzed	Analyst	Prepped	Prepped By	Specimen Wt. (g)	Analysis SOP
Pesticide LCQQQ	2639137	LCMSMS	9/26/2026 4:59:49PM	2670-10866	9/25/2026 9:05:00AM	4599	1.02	SOP1350
* Pesticides GCQQQ	2639139	GCMSMS	9/26/2026 9:45:02AM	2670-10866	9/25/2026 9:10:00AM	4599	1.02	SOP1356

Analyte	DIL	Action Limit	LOD	Results	Status
		ppb	ppb	ppb	
Abamectin	5	300	30.0	ND	Pass
Acephate	5	3000	300	ND	Pass
Acequinocyl	5	2000	200	ND	Pass
Acetamiprid	5	3000	300	ND	Pass
Aldicarb	5	100	10.0	ND	Pass
Azoxystrobin	5	3000	300	ND	Pass
Bifenazate	5	3000	300	ND	Pass
Bifenthrin	5	500	50.0	ND	Pass
Boscalid	5	3000	300	ND	Pass
Carbaryl	5	500	50.0	ND	Pass
Carbofuran	5	100	10.0	ND	Pass
Chlorantraniliprole	5	3000	300	ND	Pass
Chlormequat Chloride	5	3000	300	ND	Pass
Chlorpyrifos	5	100	10.0	ND	Pass
Clofentezine	5	500	50.0	ND	Pass
Coumaphos	5	100	10.0	ND	Pass
Cyfluthrin	5	1000	100	ND	Pass
Cypermethrin	5	1000	100	ND	Pass
Daminozide	5	100	10.0	ND	Pass
Diazinon	5	200	20.0	ND	Pass
Dichlorvos	5	100	10.0	ND	Pass
Dimethoate	5	100	10.0	ND	Pass
Dimethomorph	5	3000	300	ND	Pass
Ethoprophos	5	100	10.0	ND	Pass
Etofenprox	5	100	10.0	ND	Pass
Etoxazole	5	1500	150	ND	Pass
Fenhexamid	5	3000	300	ND	Pass
Fenoxycarb	5	100	10.0	ND	Pass
Fenpyroximate	5	2000	200	ND	Pass
Fipronil	5	100	10.0	ND	Pass
Flonicamid	5	2000	200	ND	Pass
Fludioxonil	5	3000	300	ND	Pass
Hexythiazox	5	2000	200	ND	Pass
Imazalil	5	100	10.0	ND	Pass

Analyte	DIL	Action Limi	LOD	Results	Status
		ppb	ppb	ppb	
Imidacloprid	5	3000	300	ND	Pass
Kresoxim-methyl	5	1000	100	ND	Pass
Malathion	5	2000	200	ND	Pass
Metalaxyl	5	3000	300	ND	Pass
Methiocarb	5	100	10.0	ND	Pass
Methomyl	5	100	10.0	ND	Pass
Mevinphos	5	100	10.0	ND	Pass
Myclobutanil	5	3000	300	ND	Pass
Naled	5	500	50.0	ND	Pass
Oxamyl	5	500	50.0	ND	Pass
Paclobutrazol	5	100	10.0	ND	Pass
Permethrin	5	1000	100	ND	Pass
Phosmet	5	200	20.0	ND	Pass
Piperonyl butoxide	5	3000	300	ND	Pass
Prallethrin	5	400	40.0	ND	Pass
Propiconazole	5	1000	100	ND	Pass
Propoxur	5	100	10.0	ND	Pass
Pyrethrins	5	1000	100	ND	Pass
Pyridaben	5	3000	300	ND	Pass
Spinetoram, total	5	3000	300	ND	Pass
Spinosad, total	5	3000	300	ND	Pass
Spiromesifen	5	3000	300	ND	Pass
Spirotetramat	5	3000	300	ND	Pass
Spiroxamine	5	100	10.0	ND	Pass
Tebuconazole	5	1000	100	ND	Pass
Thiacloprid	5	100	10.0	ND	Pass
Thiamethoxam	5	1000	100	ND	Pass
Trifloxystrobin	5	3000	300	ND	Pass
Captan*	5	3000	300	ND	Pass
Chlordane*	5	100	10.0	ND	Pass
Chlorfenapyr*	5	100	10.0	ND	Pass
Methyl parathion*	5	100	10.0	ND	Pass
Pentachloronitrobenzene*	5	200	20.0	ND	Pass

LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.
* - GCMSMS Pesticides

TL Laboratories
10350 Fisher Ave, Tampa, Florida 33619
813-731-2451 / www.terplifelabs.com

Brian C. Spann
Laboratory Director

PJLA Testing
Accreditation #116557

DEA Registration Number: RT0720878

The data contained therein are based on sound scientific analytical procedures and judgment. TL Laboratories strives to deliver high quality results. This report shall not be reproduced without written consent from TL Laboratories The results of this report relate only to the material or product received and analyzed. Test Results are confidential unless explicitly waived otherwise. This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISOILAC-IAF Communiqué dated April 2017)

TL LABORATORIES

Certificate of Analysis

T6I1151



JAG Alliance, LLC
Client Lic#:
83 Knight Boxx Rd Orange Park, FL
32065
JAG Alliance, LLC

Compliance for Retail: **PASS**

1000mg CBN Tincture Drop (MCT) - Mint - 1oz/ 30ml

Sample ID: T6I1151-01 CCB ID:Lot E258611546-1
Matrix: Derivative Non-inhalable

Total Sample Received: 30ml
Total Units Received: 1
Unit Weight: 30g

Date Sampled: 09/24/2026
Date Received: 09/24/2026
Date Reported: 09/29/2026

Mycotoxins — Pass

Date Prepared: 09/25/2026 09:05 Prepped By: 4599 Specimen Wt: 1.02 g
Date Analyzed: 09/25/2026 21:25 Analyzed By: 2670-10866 Instrument: LCMSMS
Lab Batch: 2639137 Prep Method: SOP 1363 Analysis Method: SOP 1350

Analyte	DIL	Action Limit	LOD	Results	Status
		ppb	ppb	ppb	
Aflatoxin B1	5	20.0	2	ND	Pass
Aflatoxin B2	5	20.0	2	ND	Pass
Aflatoxin G1	5	20.0	2	ND	Pass
Aflatoxin G2	5	20.0	2	ND	Pass
Ochratoxin A	5	20.0	2	ND	Pass

LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

Microbials — Pass

Date Prepared: 09/24/2026 15:59 Prep ID: 1093 Specimen Wt: 1.00 g
Date Analyzed: 09/29/2026 11:00 Analyst ID: 1093 Instrument: qPCR
Lab Batch: 2639112 Analysis Method: SOP1353/1364/1352

Analyte	Action Limit	LOD	Results	Status
	cfu/g	cfu/g	cfu/g	
Aspergillus Flavus	1.00	1	Absent in 1 gram	Pass
Aspergillus Fumigatus	1.00	1	Absent in 1 gram	Pass
Aspergillus Niger	1.00	1	Absent in 1 gram	Pass
Aspergillus Terreus	1.00	1	Absent in 1 gram	Pass
Salmonella	1.00	1	Absent in 1 gram	Pass
Shiga Toxin producing E. Coli	1.00	1	Absent in 1 gram	Pass
Total Yeast and Mold	100000	10000	ND	Pass

LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

Heavy Metals — Pass

Date Prepared: 09/24/2026 15:45 Prepped By: 1094 Specimen Wt: 0.54 g
Date Analyzed: 09/27/2026 00:21 Analyzed By: 10198 Instrument: ICPMS
Lab Batch: 2639103 Prep Method: SOP 1362 Analysis Method: SOP1358

Analyte	DIL	Action Limit	LOD	Results	Status
		ppb	ppb	ppb	
Arsenic	1	1500	150	ND	Pass
Cadmium	1	500	50.0	ND	Pass
Lead	1	500	50.0	ND	Pass
Mercury	1	3000	300	ND	Pass
Total				0.00	

LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

Foreign Materials — Pass

Date Prepared: 09/24/2026 16:00 Prep ID: 5046 Instrument: Visual Inspection
Date Analyzed: 09/24/2026 16:08 Analyst ID: 3780
Lab Batch: 2639108 Analysis Method: SOP1359

Analyte	Action Limit (% by wt)	Results	Status
Foreign Material	1.00	ND	Pass
Feces	0.500	ND	Pass

ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

Water Activity — Pass

Date Prepared: 09/24/2026 16:05 Prep ID: 5046 Specimen Wt: 0.16 g
Date Analyzed: 09/24/2026 16:05 Analyst ID: 3780 Instrument: Water Activity Meter
Lab Batch: 2639108 Analysis Method: SOP1355

Analyte	Action Limit	Result	Status
	aW	aW	
Water Activity	0.85	0.58	Pass

ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

TL Laboratories
10350 Fisher Ave, Tampa, Florida 33619
813-731-2451 / www.terplifelabs.com

DEA Registration Number: RT0720878

Brian C. Spann
Laboratory Director

PJLA Testing
Accreditation #116557

The data contained therein are based on sound scientific analytical procedures and judgment. TL Laboratories strives to deliver high quality results. This report shall not be reproduced without written consent from TL Laboratories The results of this report relate only to the material or product received and analyzed. Test Results are confidential unless explicitly waived otherwise. This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISOILAC-IAF Communiqué dated April 2017)

TL LABORATORIES

Certificate of Analysis

T6I1151



JAG Alliance, LLC
Client Lic#:
83 Knight Boxx Rd Orange Park, FL
32065
JAG Alliance, LLC

Compliance for Retail: **PASS**

1000mg CBN Tincture Drop (MCT) - Mint - 1oz/ 30ml

Sample ID: T6I1151-01 CCB ID:Lot E258611546-1
Matrix: Derivative Non-inhalable

Total Sample Received: 30ml
Total Units Received: 1
Unit Weight: 30g

Date Sampled: 09/24/2026
Date Received: 09/24/2026
Date Reported: 09/29/2026

Residual Solvents

Pass

Date Prepared: 09/24/2026 14:20 Prep ID: 13158 Specimen Wt: 0.05 g
Date Analyzed: 09/25/2026 04:07 Analyst ID: 13158 Instrument: GCMSMS Headspace
Lab Batch: 2639098 Analysis Method: SOP1361

Analyte	DIL	Action Limit	LOD	Results	Status
		ppm	ppm	ppm	
1,1-Dichloroethene	1	8.00	0.800	ND	Pass
1,2-Dichloroethane	1	2.00	0.200	ND	Pass
Acetone	1	5000	50.0	ND	Pass
Acetonitrile	1	410	41.0	ND	Pass
Benzene	1	1.00	0.100	ND	Pass
Butane	1	5000	500	ND	Pass
Chloroform	1	2.00	0.200	ND	Pass
Ethanol	1	5000	500	ND	Pass
Ethyl acetate	1	5000	50.0	ND	Pass
Ethyl ether	1	5000	500	ND	Pass
Ethylene oxide	1	5.00	0.500	ND	Pass
Heptane	1	5000	500	ND	Pass
Hexane	1	250	25.0	ND	Pass
Isopropyl alcohol	1	500	50.0	ND	Pass
Methanol	1	3000	300	ND	Pass
Methylene chloride	1	125	12.5	ND	Pass
Pentane	1	5000	500	ND	Pass
Propane	1	5000	500	ND	Pass
Toluene	1	890	89.0	ND	Pass
Xylenes, total	1	2170	217	ND	Pass
Trichloroethylene	1	25.0	2.50	ND	Pass

LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the Laboratory.

Terpenes Summary

Date Prepared: 09/25/2026 08:56 Prep ID: 4599 Specimen Wt: 0.11 g
Date Analyzed: 09/26/2026 09:53 Analyzed ID: 2447 Instrument: GCMS
Lab Batch: 2639135 Prep/Analysis Method: SOP1360

Analyte	Dilution	LOD	Results	
		%	%	
Menthol	1	0.0231	0.171	
3-Carene (+)-	1	0.0231	ND	
alpha Bisabolol, L	1	0.0231	ND	
alpha-Cedrene	1	0.0231	ND	
alpha-Fenchyl alcohol, (+)-	1	0.0231	ND	
alpha-Humulene	1	0.0231	ND	
alpha-Phellandrene	1	0.0231	ND	
alpha-Pinene	1	0.0231	ND	
alpha-Terpinene	1	0.0231	ND	
alpha-Terpineol	1	0.0231	ND	
beta-Myrcene	1	0.0231	ND	
beta-Ocimene	1	0.0231	ND	
beta-Pinene	1	0.0231	ND	
Borneol	1	0.0231	ND	
Camphene	1	0.0231	ND	
Camphor	1	0.0231	ND	
Caryophyllene Oxide	1	0.0231	ND	
Cedrol	1	0.0231	ND	
D-Limonene	1	0.0231	ND	
E-Caryophyllene	1	0.0231	ND	
E-Nerolidol	1	0.0231	ND	
Eucalyptol	1	0.0231	ND	
Farnesene	1	0.0231	ND	
Fenchone	1	0.0231	ND	
gamma-Terpinene	1	0.0231	ND	
Geraniol	1	0.0231	ND	
Geranyl Acetate	1	0.0231	ND	
Guaiol	1	0.0231	ND	
Isoborneol	1	0.0231	ND	
Isopulegol	1	0.0231	ND	
Linalool	1	0.0231	ND	
Nerol	1	0.0231	ND	
p-Cymene	1	0.0231	ND	
Pulegone	1	0.0231	ND	
Sabinene	1	0.0231	ND	
Sabinene hydrate	1	0.0231	ND	
Terpinolene	1	0.0231	ND	
Valencene	1	0.0231	ND	
Z-Nerolidol	1	0.0231	ND	
Total Terpenes			0.171	51.30 mg/Unit

Terpene results are provided for informational purposes only.
LOD = Limit of Detection; ND = Not Detected.
Unless otherwise stated all quality control samples performed within specifications established by the

TL Laboratories
10350 Fisher Ave, Tampa, Florida 33619
813-731-2451 / www.terplifelabs.com

Brian C. Spann
Laboratory Director

PJLA Testing
Accreditation #116557

DEA Registration Number: RT0720878

The data contained therein are based on sound scientific analytical procedures and judgment. TL Laboratories strives to deliver high quality results. This report shall not be reproduced without written consent from TL Laboratories The results of this report relate only to the material or product received and analyzed. Test Results are confidential unless explicitly waived otherwise. This laboratory is accredited in accordance with the recognised International Standard ISO/IEC 17025. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISOILAC-IAF Communiqué dated April 2017)